

International Journal of Medical Ophthalmology



E-ISSN: 2663-8274
P-ISSN: 2663-8266
www.ophthalmoljournal.com
IJMO 2023; 5(2): 65-68
Received: 02-11-2023
Accepted: 05-12-2023

Anil Kumar Yadav
Research Scholar, Department
of Optometry and Vision
Science, Opjs University
Churu, Rajasthan, India

Dr. Kapil Dev
Associate Professor,
Department of Optometry and
Vision Science, Opjs
University Churu, Rajasthan,
India

Corresponding Author:
Anil Kumar Yadav
Research Scholar, Department
of Optometry and Vision
Science, Opjs University
Churu, Rajasthan, India

Optical coherence tomography - current and future Applications

Anil Kumar Yadav and Dr. Kapil Dev

DOI: <https://doi.org/10.33545/26638266.2023.v5.i2a.187>

Abstract

Optical coherence tomography (OCT) has revolutionized the clinical practice of ophthalmology. It is a noninvasive imaging technique that provides high-resolution, cross-sectional images of the retina, retinal nerve fiber layer and the optic nerve head. This review discusses the present applications of the commercially available spectral-domain OCT (SD-OCT) systems in the diagnosis and management of retinal diseases, with particular emphasis on choroidal imaging. Future directions of OCT technology and their potential clinical uses are discussed. Analysis of the choroidal thickness in healthy eyes and disease states such as age-related macular degeneration, central serous chorioretinopathy, diabetic retinopathy and inherited retinal dystrophies has been successfully achieved using SD-OCT devices with software improvements. Future OCT innovations such as longer-wavelength OCT systems including the swept-source technology, along with Doppler OCT and en-face imaging, may improve the detection of subtle microstructural changes in chorioretinal diseases by improving imaging of the choroid. Advances in OCT technology provide for better understanding of pathogenesis, improved monitoring of progression and assistance in quantifying response to treatment modalities in diseases of the posterior segment of the eye. Further improvements in both hardware and software technologies should further advance the clinician's ability to assess and manage chorioretinal diseases.

Keywords: Applications of optical coherence tomography, chorioretinal diseases, retina, spectral-domain optical coherence tomography, swept-source optical coherence tomography

Introduction

Optical coherence tomography (OCT) has evolved over the past decade as one of the most important ancillary tests in ophthalmic practice. It is a noninvasive imaging technique and provides high-resolution, cross-sectional images of the retina, the retinal nerve fiber layer (RNFL) and the optic nerve head. With axial resolution in the 5-7 μm range, it provides close to an in-vivo 'optical biopsy' of the retina. OCT employs light from a broadband light source, which is divided into a reference and a sample beam, to obtain a reflectivity versus depth profile of the retina. The light waves that are backscattered from the retina, interfere with the reference beam, and this interference pattern is used to measure the light echoes versus the depth profile of the tissue *in vivo*.

At its advent, time-domain detection was the technique employed by commercially available OCT systems such as the Stratus OCT (Carl Zeiss Meditec, Inc, Dublin, CA). Time-domain OCT (TD-OCT) systems featured scan rates of 400 A-scans per second with an axial resolution of 8-10 μm in tissue. In 2006, the first commercially available spectral-domain (Fourier domain) OCT (SD-OCT) system was introduced. SD-OCT employs detection of the light echoes simultaneously by measuring the interference spectrum, using an interferometer with a high-speed spectrometer. This technique achieves scan rates of 20 000-52 000 A-scans per second and a resolution of 5-7 μm in tissue.

Although OCT is used extensively for clinical decision making and monitoring of many posterior segment diseases based on macular, optic nerve and RNFL imaging, until recently, the choroid was not able to be clearly imaged with this technique. New innovations in SD-OCT hardware and software now allow for accurate choroidal thickness measurements. In addition, choroidal morphological changes on OCT are now being appreciated. As a result, choroidal imaging is an emerging area of research.

The choroid cannot be well visualized using the Stratus OCT, as the retinal pigment epithelium (RPE) is highly light scattering, resulting in attenuation of the relatively weak reflection signal from the choroid. In addition, because of the relatively low signal-to-noise ratio of TD-OCT, the signal and image information from the deeper layers of the choroid is not of high enough quality to see precise morphological details.

Also, the pixel density of TD-OCT, which is limited by the number of axial scans in the OCT image, makes visualization of the choroid difficult.

SD-OCT systems can image the choroid, however, using techniques such as image averaging and enhanced depth imaging (EDI). Image averaging involves obtaining multiple B-scans from the same retinal location that are then averaged together to increase the signal-to-noise ratio, typically in proportion to the square root of the number of images averaged. When multiple images are averaged, the software reduces the 'speckle.' This sharpens the continuity and enhances the retinal and choroidal features. Along with image averaging, EDI involves setting the choroid adjacent to the zero delay line, which allows enhanced visualization of choroid up to the sclera, by taking advantage of the sensitivity roll-off characteristic of SD-OCT systems.

Apart from the commercially available systems, prototype OCT systems have contributed to an ever-growing body of research studies in this field. These include, but are not limited to, the ultra-high-resolution OCT (UHR-OCT), SD-OCT systems employing a longer-wavelength light source permitting deeper tissue penetration, and swept-source OCT (SS-OCT) systems. UHR-OCT uses broadband light sources to achieve 3 μm resolution in tissue. SS-OCT uses another form of Fourier domain detection to measure light echoes. It employs a tunable frequency swept laser light source, which sequentially emits various frequencies in time, and the interference spectrum is measured by photodetectors instead of a spectrometer. This increases the signal quality in deep tissue by elimination of the sensitivity of a spectrometer to higher frequency modulation as with SD-OCT, thereby improving the visualization of the choroid.

This review discusses the present applications of the various commercially available SD-OCT systems in the diagnosis and management of retinal diseases, with particular emphasis on choroidal imaging. Further, future directions of OCT technology, specifically the innovations in OCT technology with the research prototype OCT systems, their potential clinical uses, and benefits for choroidal imaging are discussed.

Present Applications of Optical Coherence Tomography

At present, OCT imaging is used extensively for imaging the macula, optic nerve and RNFL, and aids in analyzing the morphology and quantifying changes in various disease states. For example, the automated retinal thickness measurements generated by the SD-OCT systems are used clinically for the monitoring progression of diseases such as wet age-related macular degeneration (AMD) and macular edema from various causes including diabetes and retinal vein occlusion. The ability to detect fluid within the retina and the thickness alterations induced by this fluid helps direct clinical decisions regarding treatment. The diagnosis of macular hole and its differentiation between lamellar holes and pseudomacular holes have become straightforward using OCT. In addition, the size and configuration of macular holes, determined by OCT, correlates well with the functional and anatomic outcomes following surgical intervention. Evaluation of the vitreoretinal interface using OCT is important in the evaluation and treatment of diseases of the vitreomacular interface such as epiretinal membranes and vitreomacular traction. In addition, optic disc morphology and RNFL thickness measurements, using OCT, monitor the

progression and helps quantitatively assess the treatment response in patients with glaucoma.

Choroidal imaging in healthy eyes

Most commercially available SD-OCT systems can be used to evaluate choroidal thickness. The method for choroidal thickness analysis involves manual measurements perpendicularly from the outer edge of the hyperreflective RPE to the inner sclera (choroid-sclera junction) using the respective software within the system. The choroid in healthy eyes is thickest subfoveally and thins nasally more than temporally. In addition, a negative correlation exists between choroidal thickness and age. Margolis *et al.* reported the mean subfoveal choroidal thickness of $287 \pm 76 \mu\text{m}$ using the Spectralis OCT and $272 \pm 81 \mu\text{m}$ using the Cirrus OCT device in 54 and 34 healthy eyes, respectively. A recent investigation analyzed the reproducibility of choroidal thickness measurements, showing a good intervisit and interobserver correlation of measurements ($r = 0.89$ and $r = 0.97$ respectively, $P < 0.05$). In addition, among the three commercially available SD-OCT systems, a good reproducibility of choroidal thickness measurements has been reported ($P < 0.0001$).

Presently, SD-OCT systems are being used extensively for evaluation of the choroid in many posterior segment diseases. OCT applications of some of the common diseases are discussed.

Age-related macular degeneration

Age-related macular degeneration (AMD) is a leading cause of visual impairment in patients over 60 years of age in developed countries. The evaluation and management of AMD utilize several investigation modalities, but advancements in OCT technology have significantly contributed to better understanding of the disease, and have helped with monitoring the progression and therapeutic response of corticosteroids and anti-vascular endothelial growth factor (VEGF) agents in wet (neovascular) AMD. Choroidal involvement in the pathogenesis of AMD has earned great interest recently. Manjunath *et al.* reported a mean subfoveal choroidal thickness of $194.6 \pm 88.4 \mu\text{m}$ in 40 eyes and $213.4 \pm 92.2 \mu\text{m}$ in 17 eyes with wet and dry AMD, respectively, when compared with $272 \pm 81 \mu\text{m}$ in age-matched healthy eyes. In addition, this study also reported an inverse correlation of age with choroidal thickness in eyes with dry AMD ($r = -0.7$, $P = 0.002$). This was further supported by Spaide, who described a distinct entity known as age-related choroidal atrophy, suggesting the contribution of both, increasing age as well as the choroidal vasculature, in the pathogenesis of AMD. Other diseases of the choroidal circulation, including central serous chorioretinopathy (CSCR) and polypoidal choroidal vasculopathy (PCV), demonstrate a thicker choroid when compared with healthy eyes and those with wet AMD. Thus, the assessment of choroidal thickness using SD-OCT may assist in differentiating neovascular AMD from CSCR and/or PCV.

Central serous chorioretinopathy

CSCR is a disease characterized by an exudative/serous detachment of the neurosensory retina from the RPE. Conventional investigations such as indocyanine green (ICG) angiography and fluorescein angiography show a generalized disruption of the choroidal vasculature in eyes affected with CSCR, with diffuse hyperpermeability. On

OCT, an elevation of the neurosensory retina from the RPE, with an optically empty space in between, is observed. Recent studies using three different SD-OCT devices reported a significant increase in the thickness of the choroid in eyes affected with acute CSCR. In 28 eyes of 19 patients with CSCR, the mean choroidal thickness was $505 \pm 124 \mu\text{m}$ on Spectralis OCT. This was 85% greater than that of age-matched healthy eyes ($P < 0.001$). This finding suggests that the pathophysiology of CSCR may involve an increase in the hydrostatic pressure in choroidal vessels. Generally, CSCR resolves spontaneously within 3-6 months, but in some patients subretinal fluid can persist for many months, leading to permanent visual dysfunction. This chronic form of CSCR may require intervention with treatments such as laser photocoagulation and photodynamic therapy (PDT). Recently, an investigation revealed a significant reduction in choroidal thickness, from $389 \pm 106 \mu\text{m}$ at baseline to $330 \pm 103 \mu\text{m}$ ($P < 0.001$), 4 weeks following PDT in eyes affected with CSCR. This, however, was not true for eyes that were treated with laser photocoagulation 4 weeks after treatment ($345 \pm 127 \mu\text{m}$ versus $340 \pm 124 \mu\text{m}$, $P = 0.2$). Such a reduction in choroidal thickness following PDT for CSCR has been demonstrated in another more recent investigation as well given the widespread use of PDT for the treatment of CSCR, analysis of choroidal thickness using SD-OCT may be a parameter to assess for disease activity following treatment. In addition, evaluation of choroidal thickness may assist in differentiating CSCR from other causes of exudative/serous retinal detachment.

Diabetic retinopathy

Diabetic retinopathy is the leading cause of visual impairment in working-age adults worldwide. The pathogenesis and clinical features of diabetic retinopathy are primarily attributed to retinal vascular damage. However, studies on vascular filling of the choroidal vessels suggest that choroidal angiopathy may also be involved. Delayed choroidal vascular filling, which manifests as choroidal hypofluorescence on ICG, is observed in eyes affected with diabetic retinopathy, and correlates with the severity of disease. Using SD-OCT, analysis of choroidal thickness in eyes affected with diabetic retinopathy has been performed. Regatieri *et al.* reported a significant thinning of the choroid in eyes with proliferative diabetic retinopathy and diabetic macular edema, but not in eyes with nonproliferative diabetic retinopathy, when compared with age-matched healthy eyes. Esmaeelpour *et al.* analyzed the subfoveal choroidal thickness in eyes with different clinical features of diabetic retinopathy, and found that eyes with microaneurysms, hard exudates and macular edema had a mean subfoveal choroidal thickness of $208 \pm 49 \mu\text{m}$, $205 \pm 54 \mu\text{m}$ and $211 \pm 76 \mu\text{m}$, respectively, when compared with $327 \pm 74 \mu\text{m}$ in healthy eyes ($P < 0.001$). This suggests that choroidal angiopathy may be related to disease severity, and choroidal thickness analysis using SD-OCT may be an important parameter to assess for the severity of diabetic retinopathy.

Evaluation of various features in the retina of patients with retinal dystrophies, in particular retinitis pigmentosa, has been investigated using SD-OCT. However, the choroid has only recently been evaluated in these diseases using SD-OCT. Scanning Doppler flowmetry studies show that choroidal blood flow is diminished, and correlates with the cone function in eyes affected with retinitis pigmentosa.

Yeoh *et al.* first described the changes in the structure and thickness of the choroid in retinal dystrophies other than retinitis pigmentosa, and reported focal and/or diffuse choroidal thinning in eyes with severe disease. Recently, Dhoot *et al.* demonstrated a mean choroidal thickness of $245.6 \pm 103 \mu\text{m}$ in 21 patients with retinitis pigmentosa, in comparison with $337.8 \pm 109 \mu\text{m}$ in healthy eyes ($P < 0.0001$) using Spectralis EDI OCT. These studies provide a new insight regarding the involvement of choroid in inherited retinal dystrophies.

Future Directions

With the growing evidence regarding the involvement of choroid in retinal diseases, it has become increasingly important to visualize the choroid in anatomic detail. Up until recently, the choroid could not be visualized effectively. Recent innovations in technology have helped with overcoming this limitation. Further ongoing advancements are expected to provide an even better understanding of choroidal involvement in retinal diseases using various techniques.

Longer wavelength and swept source technology

An accurate assessment of the choroid using OCT requires that the choroid be visualized up to the choroid-sclera interface. Studies using Cirrus OCT system report a clear visualization of the choroid-scleral interface in 70-75% of healthy and diseased eyes. To our knowledge, the percentage of eyes that have a clearly delineated choroid-sclera interface has not been reported with Spectralis and RTVue OCT systems. For adequate analysis of choroidal thickness and volume in healthy and diseased states, the clarity of the choroid-sclera interface is imperative. This can be achieved by increasing the depth of tissue penetration using a longer wavelength of incident light centered near 1050 nm, so that attenuation from scattering can be reduced. Prototype OCT systems using longer wavelength have demonstrated an enhanced visualization of the choroid, also through opaque media. SS-OCT technology is also expected to improve the visualization of the choroid. Because swept laser light sources can rapidly sweep the required frequencies, the acquisition of scans is much faster in SS-OCT, when compared with the SD-OCT systems. The SS-OCT systems have axial scan rates of up to 100 000-236 000 A-scans per second, which is five to 10 times that of the SDOCT systems, and can achieve 11 μm axial resolutions in tissue. Because data can be acquired much faster, volumetric assessment of the choroid is also feasible.

As longer-wavelength OCT systems including SS-OCT become available, the visualization of choroid-sclera interface is expected to improve. This is important in diseases such as CSCR, where the choroid is thicker than normal, and thus difficult to evaluate across its entire width. In addition, volumetric analysis of the choroid, as well as that of the various pathological features such as choroidal neovascularization and subretinal/intraretinal fluid, may be possible. Such a volumetric analysis is expected to help with monitoring the progression of diseases such as wet AMD, CSCR and diabetic retinopathy, as well as assessment of the response to treatments such as anti-VEGF agents, laser photocoagulation and PDT.

Doppler optical coherence tomography

In contrast to ICG and fluorescein angiography, which are two-dimensional investigations for blood flow analysis,

Doppler OCT is a promising technology, in that it is depth resolved, such that precise location of vascular abnormalities can be localized using cross-sectional imaging. Doppler OCT can evaluate blood flow and volume of retinal and choroidal vasculature, highlight vessels where the flow is present and evaluate abnormalities in retinal and choroidal vasculature. Given the evidence of choroidal angiopathy in various retinal diseases, this technology promises to help with monitoring of chorioretinal diseases, in particular wet AMD. It is also expected to aid in the differentiation among diseases such as wet AMD, CSCR and PCV.

En-face imaging

Software modifications, improvements and efficient processing of data are important for effective evaluation of changes in retina and choroid in posterior segment diseases. One of the advancements, known as en-face imaging, allows the clinician to visualize three-dimensional data in a fundus projection. Using this technique, particular retinal and/or choroidal layers at a given depth are projected onto an en-face view. Although cross-sectional images (B-scans) have helped delineate pathological features in retinal diseases, as such, microstructural changes and morphology of the retinal and choroidal vasculature are hard to evaluate using B-scans. This is expected to improve as en-face imaging provides further detail about the subtle pathological features in the retina and choroid in diseased states. In addition, the involvement of the specific vascular layers of the choroid in different diseases such as AMD, CSCR, diabetic retinopathy and inherited retinal dystrophies is expected to delineate in further detail using this technique.

Conclusion

OCT technology provides for enhancement of the understanding, monitoring progression and response to various treatment modalities employed in chorioretinal diseases. As such, these advancements have revolutionized ophthalmic practice over the last decade. Further innovations in both hardware and software technologies are expected to aid in the assessment of chorioretinal diseases in more detail.

References

- Huang D, Swanson EA, Lin CP, *et al.* Optical coherence tomography. *Science*. 1991;254:1178-1181.
- Sull AC, Vuong LN, Price LL, *et al.* Comparison of spectral/Fourier domain optical coherence tomography instruments for assessment of normal macular thickness. *Retina*. 2010;30:235-245.
- de Boer JF, Cense B, Park BH, *et al.* Improved signal-to-noise ratio in spectral-domain compared with time-domain optical coherence tomography. *Opt Lett*. 2003;28:2067-2069.
- Leitgeb R, Hitzenberger C, Fercher A. Performance of Fourier domain vs. time domain optical coherence tomography. *Opt Express*. 2003;11:889-894.
- Spaide RF, Koizumi H, Pozzoni MC. Enhanced depth imaging spectral-domain optical coherence tomography. *Am J Ophthalmol*. 2008;146:496-500.
- Manjunath V, Taha M, Fujimoto JG, Duker JS. Choroidal thickness in normal eyes measured using Cirrus HD optical coherence tomography. *Am J Ophthalmol*. 2010;150:325e1-329e1.
- Imamura Y, Fujiwara T, Margolis R, Spaide RF.

Enhanced depth imaging optical coherence tomography of the choroid in central serous chorioretinopathy. *Retina*. 2009;29:1469-1473.

- Branchini L, Regatieri CV, Flores-Moreno I, *et al.* Reproducibility of choroidal thickness measurements across three spectral domain OCT systems. *Ophthalmology*. 2012;119:119-123.
- Regatieri CV, Branchini L, Carmody J, *et al.* Choroidal thickness in patients with diabetic retinopathy analyzed by spectral-domain optical coherence tomography. *Retina*. 2012;32:563-568.
- Manjunath V, Goren J, Fujimoto JG, Duker JS. Analysis of choroidal thickness in age-related macular degeneration using spectral-domain optical coherence tomography. *Am J Ophthalmol*. 2011;152:663-668.
- Ho J, Witkin AJ, Liu J, *et al.* Documentation of intraretinal retinal pigment epithelium migration via high-speed ultrahigh-resolution optical coherence tomography. *Ophthalmology*. 2011;118:687-693.
- Grulkowski I, Liu JJ, Potsaid B, *et al.* Retinal, anterior segment and full eye imaging using ultrahigh speed swept source OCT with vertical-cavity surface emitting lasers. *Biomed Opt Express*. 2012;3:2733-2751.
- Unterhuber B, Povazay B, Hermann H. *In vivo* retinal optical coherence tomography at 1040 nm-enhanced penetration into the choroid. *Opt Express*. 2005;13:3252-3258.
- Chinn SR, Swanson EA, Fujimoto JG. Optical coherence tomography using a frequency-tunable optical source. *Opt Lett*. 1997;22:340-342.
- de Boer JF, Milner TE, van Gemert MJ, Nelson JS. Two-dimensional birefringence imaging in biological tissue by polarization-sensitive optical coherence tomography. *Opt Lett*. 1997;22:934-936.
- Drexler W, Fujimoto JG. State-of-the-art retinal optical coherence tomography. *Prog Retin Eye Res*. 2008;27:45-88.

How to Cite This Article

Yadav AK and Dr. Dev K. Optical coherence tomography - current and future Applications. *International Journal of Medical Ophthalmology*. 2023; 5(2): 65-68

Creative Commons (CC) License

This is an open access journal, and articles are distributed under the terms of the Creative Commons Attribution-NonCommercial-ShareAlike 4.0 International (CC BY-NC-SA 4.0) License, which allows others to remix, tweak, and build upon the work non-commercially, as long as appropriate credit is given and the new creations are licensed under the identical terms.